When To Replace Your Breast Pump Parts



Flanges: every 6months; check often for damage, replace as needed

Tubing: every 6 months or when damaged, moisture inside or slipping off of pump

Valves: 1-3months

Valve Membranes: 2-8weeks

Backflow

Protectors: 3-6months

These ranges are dependent

on usage

Contact for help with pump parts if needed





